



## C-Star Calibration

Date **November 11, 2022** S/N# **CST-2405** Pathlength **25 cm**

---

	Analog output	Digital output	
$V_{\text{dark}}$	<b>0.016 V</b>	<b>0 counts</b>	
$V_{\text{air}}$	<b>4.792 V</b>	<b>15661 counts</b>	
$V_{\text{ref}}$	<b>4.699 V</b>	<b>15356 counts</b>	
Temperature of calibration water			<b>19.6 °C</b>
Ambient temperature during calibration			<b>21.5 °C</b>

---

Relationship of transmittance ( $Tr$ ) to beam attenuation coefficient ( $c$ ), and pathlength ( $x$ , in meters):  $Tr = e^{-cx}$

To determine beam transmittance:  $Tr = (V_{\text{sig}} - V_{\text{dark}}) / (V_{\text{ref}} - V_{\text{dark}})$

To determine beam attenuation coefficient:  $c = -1/x * \ln(Tr)$

$V_{\text{dark}}$  Meter output with the beam blocked. This is the offset.

$V_{\text{air}}$  Meter output in air with a clear beam path.

$V_{\text{ref}}$  Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain  $V_{\text{ref}}$ .

Ambient temperature: meter temperature in air during the calibration.

$V_{\text{sig}}$  Measured signal output of meter.